

MAR 29 2011

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130176</u> ✓	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>DARYL RICE</u>	ID# <u>2088</u> Date <u>03/19/11</u> ✓
A Agency <u>SITKA POLICE DEPT.</u> Phone # <u>747-3245</u>	
Instrument Location <u>304 LAKE STREET</u>	
B Alco S/N <u>X301572</u> ✓ Target Value <u>.078</u> ✓ High Pressure <u>500</u>	
Alco Test Values <u>.088</u> ✓ <u>.078</u> ✓	
1st Alco 2nd Alco	
Signature <u>Daryl Rice</u>	
(OVER)	

(Do Not write in the area below)

CSE
3/31/11


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 1st day of APRIL, 2011 DN

Daniel L. Nicholas
(Notary Seal Stamp)
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

DIAGNOSTIC CHECK

ALCO S/N: X301572

— BREATH ANALYSIS —

.078 ADJUSTED FOR 30.17 in	.078	03:49
ALCO TARGET	.000	03:50
BLANK TEST	.000	03:50
INTERNAL STANDARD	VERIFIED	
ALCO TV 30.17 in	.004	03:51
BLANK TEST	.000	03:51
SUBJECT SAMPLE	.000	03:52
BLANK TEST	.000	03:53
ALCO TV 30.17 in	.078	03:53
BLANK TEST	.000	03:54

— 11 —

1. The first step is to identify the problem. In this case, the problem is that the company is not meeting its sales targets.

PRINTER TEST

```
"#3."(#+,-.0123456789::(<=>0ABCEFG  
HIMNKPQRSTUWXYZ)-1"':code=gnjklmno  
pqrstuwxzy(')-&
```